

MIDORI Green Xtra Agarose Tablets

Cat. No. AG12

Description

MIDORI Green Xtra Agarose Tablet contains a pre-determined amount of standard melting point agarose and MIDORI Green Advance stain.

Midori Green Xtra DNA stain emits green fluorescence when bound to DNA or RNA. It was developed to work with Blue and Blue/Green LED Illuminators (like the FastGene® LED Illuminator or FastGene® LED Transilluminators).

The purity of the agarose leads to an excellent transparency and a low background, important to obtain sharp and well-defined DNA/RNA bands with the highest sensitivity in the low molecular weight range.

Safety

Caution when using hot, viscous solutions! Use suitable safety gear and open bottle gently to avoid accidents.

MIDORI Green Xtra DNA stain is non-carcinogenic and according to the Ames test it causes significantly fewer mutations than Ethidium bromide. It can irritate skin and eyes. Please wear gloves while handling. A detailed safety report can be downloaded at www.nippongenetics.eu.

Specifications

- Melting point: 88 ± 1,5 °C
- Separation range: 100 bp to >30 kb
- Product size: 100 tablets (0,5 g each)

Applications

- Ideal for routine DNA and RNA gel electrophoresis and blotting assays.
- Convenient tablet format - no complicated messy weighing required.
- Fast dissolving protocol.

NIPPON Genetics EUROPE GmbH

 +49 2421 554960  info@nippongenetics.eu
 +49 2421 55496 11  www.nippongenetics.eu

Protocol

1. Use the bottle or flask that is at least 3 times the volume of the solution being prepared.
2. Add an appropriate number of agarose tablets in the running buffer! See the table below to achieve the needed gel percentage.

Gel %	1 tablet	2 tablets
1%	50 ml	100 ml
1,5 %	33 ml	66 ml
2 %	25 ml	50 ml

3. Soak the tablet in a **running buffer** for 3 minutes (or until it is dissolved) before heating. For tablet dissolving use running buffer which is at **RT**.
4. Heat the solution until it is clear and visually all the particles are dissolved.
5. Cool the gel to 60-70°C.
6. Cast the gel into the gel tray.
7. Run the gel in used running buffer.
8. Detect the bands under Blue or Blue/Green LED. UV light is not recommended.

Storage

Storage and shipping at RT. Protected from light.

Ordering information

Cat. No.	Pack Size
AG12	100 pcs
AG12s	2 pcs